

Temperature Tester 351

Instruction Manual



Test Products International, Inc.

9615 SW Allen Blvd., Ste. 104
Beaverton, OR USA 97005
503-520-9197 • Fax: 503-520-1225
info@tpi-thevalueleader

Test Products International, Ltd.

342 Bronte Road South, Unit 9
Milton, Ontario L9T5B7
Canada
905-693-8558 • Fax: 905-693-0888

Test Products International UK Ltd

Longley House
East Park
Crawley, West Sussex RH10 6AP England
Tel: 01293 561212 • Fax: 01293 813465

EC Declaration of Conformity

This is to certify that models 352/353/354 conform to the protection requirements of the council directive 89/336/EEC, in the approximation of laws of the member states relating to Electromagnetic compatibility by application of the following standards:

EN 50081-1	1992 Emissions Standard
EN 50082-1	1992 Immunity Standard

To ensure conformity with these standards, this instrument must be operated in accordance with the instructions and specifications given in this manual.



WARNING: Please follow manufacturers test procedures whenever possible. Make sure probe tip and shaft are thoroughly cleaned before inserting it into any food product.

GENERAL GUIDELINES

ALWAYS

- Test the 340 before using it to make sure it is operating properly.
- Inspect temperature probes before using to make sure there are no breaks or shorts.
- Double check all connections before testing.

NEVER

- Insert the probe into an electrically charged medium.
- Touch the probe tip and shaft after it has been inserted in hot material.

Features and Benefits

- Water Resistant** Operates in any environment.
- Auto Power Off** Instrument automatically shuts off after 30 minutes of inactivity.

Specifications

Range with Thermistor Probe

Range	Accuracy
-40° to 220°F	±1° (32° to 158° F)

General Specifications

Power Supply	9 Volt Battery
Battery Life	200 hrs. Typical
Size (HxLxW)	40mm x 190mm x 65mm (1.5" x 7.25" x 2.5")
Weight	315g (11.1 oz)

1. Controls and Functions:

Push Buttons

- ① Turns the 351 on and off.

2. Step by Step Procedures

- a. Connect the temperature probe to the 351.
- b. Turn the 351 on.
- c. Insert the tip of the probe into the substance being measured.
- d. Read the temperature on the LCD, allowing approximately 30 seconds for the reading to stabilize.

Standard Accessories

Part No.

9V Battery

A009

Test Lead Set

A040

Rubber Boot

A303

Optional Accessories	Part No.
Surface Probe	CX13B
Air Probe	GX15B
3' Extension Lead	EX11B
-18 C or -0.4 F test cap	VX11B
0 C or 32 F test cap	VX12B
70 C or 158 F test cap	VX13B
Soft Pouch	A255

MAINTENANCE

- Battery Replacement:** The 351 will display battery symbol when the internal 9 Volt battery needs replacement. Replace battery as follows:
 - Disconnect and remove all temperature probes from 351.
 - Loosen the 4 screws from the back of the 351.
 - Remove the back cover.
 - Remove old battery and replace it with new battery.
 - Reassemble instrument in reverse order from above.
- Cleaning your 351**

Use a mild detergent and slightly damp cloth to clean the surfaces of 351.